

Del Mar Photonics

4119 Twilight Ridge, San Diego, CA 92130 tel: (858) 876-3133, (858) 630-2376

CERTIFICATE OF CONFORMITY

Certificate date: May 19, 2010

Customer	Lawrence Livermore National Laboratory
Order No.	U1456696 Credit Card Order
Order date	March 8, 2010
Quotation reference	100304/1230
Description of optical component	Triggered spark gap
Drawing #	--
Part Number /CAT# /Stock code#	RU-69 triggered spark gap
Batch# or Lot#	RU-69 triggered spark gap
Qty in this delivery	12 pieces

I. Material	Lot	RU#15-7G/25
--------------------	------------	-------------

II. Polishing parameters	Specification		Actual data
	Nominal	Tolerance	
Minimum Anode Voltage, kV	2	--	meets
Maximum Anode Voltage, kV	4	--	meets
Commutated energy, J	≤ 4	--	meets
Pulse current, kA	≤ 4	--	meets
Pulse Frequency, Hz.	≤ 10	--	meets
Diameter, mm	55	--	meets
Height, mm	36	--	meets

III. Coating parameters	Specification		Actual data (measured/ extrapolated)
	Nominal	Tolerance	
--	--	--	--

IV. Notes
--



 Contact: ALEXANDER SMIRNOV
 Tel (858) 876-3133, fax (858) 630-2376
 Email: alex.smirnov@dmphotonics.com
 WEB: <http://www.dmphotonics.com/>